



8 Channel Vibrating Wire Sensor Interface

User Guide & Installation Manual

Version 1.11

Last updated 24/09/2015

WARRANTY

Keynes Controls Ltd warrants its products to be free of defects in materials and workmanship, under normal use and service for a period of 12 months from the date of purchase. If the unit should malfunction, it must be returned to Keynes Controls for evaluation, freight prepaid. Upon examination by Keynes Controls Ltd, if the unit is found to be defective, it will be repaired or replaced at no charge.

However the WARRANTY is VOID if the unit shows evidence of having been tampered with or shows evidence of being damaged as a result of excessive corrosion or current, heat, moisture or vibration., improper specification misuse outside of the companies control.

Components which wear or become damaged by misuse are not warranted. This includes batteries, fuses and connectors.

Models VibWire-108-SDI12 and VibWire-108-485 are fully integrated into the Keynes Controls Free Q-LOG Data Acquisition & Display Software. Copies of this software can be downloaded from the company web site.

Release Information

This manual refers to products sold and supplied after August 2015.

Introduction

The following document is the User Manual and installation guide for the VibWire-108 range of instruments.

The User is expected to have some prior knowledge of the SDI-12, RS-485 or Modbus network and protocols since this manual is not meant as a teaching aid for network applications.

The VibWire-108 family of vibrating wire sensor interfaces have been designed to interface vibrating wire sensors from any manufacture to a data logger, PC data acquisition system, or SCADA applications.

The principle operating feature of the VibWire-108 is its ability to accurately measure and report the vibrating wire sensor frequency. The instrument uses an auto-resonance technique to energise the sensor coil and adjusts the ping rate automatically to follow the sensor operation. The auto-resonance feature enables the sensors to be automatically configured as the instrument adjusts the coil excitation to maximise the signal to noise ratio without excessive strain put to the sensor coil.



Model VibWire-108 SDI-12



Model VibWire-108-485

Hardware Options

VibWire-108-RS485	with RS-485 network option
VibWire-108-SDI12	with SDI-12 network option
VibWire-108-Modbus	with RS-485 Modbus option
VibWire-108-Analog	with analogue output option

Dynamic & Static Measurement Applications

The VibWire-108 is ideally suited to static measurement applications. In applications requiring samples rates of 1 - 10 Samples/ Sec then a new keynes Controls product the VibWire-101-RS232 will be required. See

http://www.aquabat.net/HighSpeedVW/high-speed-vw101v1.html

Configuration

For SDI-12, RS485 and Modbus network devices the frequency input configuration settings for each of the VW sensor connected to the devices are automatically assigned.

Only the model VibWire-108-Analog instrument of the VW-108 range requires any VW sensor frequency configuration and this is this is only when the analogue output representation of the input signal is being assigned.

SI Units

The VibWire-108 can be set-up to supply results directly in units of Hz, Digits (Hz²), and Engineering units. The vibrating wire sensor engineering unit conversion is carried out using the industry standard quadratic equation expansion. The vibrating wire

FEATURES

- 8 x 4 Wire Vibrating Wire Sensor Inputs
- Resolves the VW signal to less than 0.001 Hz (industry standard 0.1 Hz)
- Gas Discharge Tube Sensor Protection
- Real-time Frequency Display 5 digit
- Audible Output
- Auto Resonance VW Excitation
- SDI-12 Network Support
- Firmware Upgrade facility
- Automatic VW sensor configuration
- Digital communications to remove noise sources and errors.
- Simplified configuration and data logger support.
- Output Frequency, Digits, SI Units, Temp Deg C
- Steinhart-Hart Thermistor Linearisation Support
- Integrated Polynomial Linearisation Quadratic Support direct from VW sensor calibration data sheet.

Field Operations

All of the VibWire-108 family of interfaces contain a real-time 5 digit, 7 segment LED display that can be used to show real-time sensor frequencies for the vibrating wire sensors, and to configure the most commonly used features of the instrument. This feature is useful when configuring and testing sensors in the field.

The VibWire-108 automatically resets to the network operation after a time-out period of 10 minutes and so prevents a User leaving in the wrong mode of operation. This feature ensures the instrument is always ready for operation and is useful for widely distributed applications and systems deployed in hard to reach locations.

Terminal Port

The VibWire-108 supports a terminal port configuration and upgrade facility. The terminal port can be used by any industry standard terminal emulator software such as the Microsoft Hyper-terminal or Token-2. The terminal port enables the vibrating wire sensor calibration factors to be stored easily into the instrument.

Fully Integrated Data Recording Solutions

The VibWire-108 can be connected to any suitable third party data logger or communication system supporting SDI-12, RS-485 or Modbus operations. Simple industry standard commands are used to make a reading and acquire data.

The Modbus network protocol is supported for easy integration into SCADA applications.

Q-LOG

The VibWire-108 is fully integrated into the free Keynes Controls Q-LOG data recording and display software. The Q-LOG application software enables the VibWire-108 to be used to create PC based data recording and display solutions with little or no programming experience.

The Q-Log software can be downloaded for free

http://www.aquabat.net/QLOGFree/qlogv2.html

Photo VibWire-108

1.0 Care & Maintenance

The VibWire-108 family of products have been designed for long term operation and so will operate reliably for many years as long as the instrument is not mis-used and operated as shown in the manual..

Step 1

Remove any signal cables and terminal blocks from the instrument.

Step 2

Clean the 4 and 5 way plug and sockets using ionised water to remove the buildup of any dirt or foreign bodies that build up on the termination pins. It is essential to remove any grease that can cause corrosion to the pins.

Step 3

Allow the sockets to dry out before connecting any signal cables.

Description	
Operating temperature	−10 to 60 °C
Storage temperature	−10 to 85 °C
Operating humidity	10 to 90% RH, non-condensing
Storage humidity	5 to 95% RH, non-condensing

2.0 Default Factory Settings

All instruments are set forNo of Channels = 8Temp = 8Default ID = 0Models VibWire-108-SDI12, VibWire-108-RS485, VibWire-108-Modbus

3.0 Required Software

The VibWire-108 requires a terminal software package supporting VT100 emulation only.

Recommended software

Microsoft Hyper-terminal Token2

3.1 Q-Log Software

The Q-Log data acquisition and display software operates with all of the USB-SDI12 and USB-RS5485 media converters

Q-Log enables the VibWire-108 to operate with a PC or laptop and give the User access to the data in a familiar Windows environment

When used with a USB-SDI12-Pro or USB-RS485-Pro media converters, the VibWire-108 can directly powered from the PC/Laptop USB port.

Download at http://www.aquabat.net/QLOGFree/qlogv2.html

Warranty Information

The information in this document is subject to change without notice. Keynes Controls Ltd . has made a reasonable effort to be sure that the information contained herein is current and accurate as of the date of publication.

Keynes Controls Ltd. makes no warranty of any kind with regard to this material, including, but not limited to, its fitness for a particular application. Keynes Controls Ltd will not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

In no event shall Keynes Controls Ltd . be liable for any claim for direct, incidental, or consequential damages arising out of, or in connection with, the sale, manufacture, delivery, or use of any product.



The keyboard controls and menu system remains the same between models



4.1 Circuit Board Features

5.0 SDI-12 Serial Network Connection

The following page describes connecting the VibWire-108 to an SDI-12 network.



assigned.

5.4

Use the same network connection for any 3rd party logger

5.1 **SDI-12 Network Operation**

12 DC

The SDI-12 multi-drop network requires only 3 wires to be powered on. Upon receipt of the correct start measurement connected between instruments for the communication of data. command then then a scan of the sensors is made and data is This ensures that the installation and use of the SDI-12 network sent in turn to the data recorder. After all data is received the the is a very simple operation. The VibWire-108 is powered by the instruments are all powered off. SDI-12 network +12V and 0V supply operations. The SDI-12 network only goes active during a measurement operation and The image below shows how a simple multi-instrument system is switched off at any other time. The SDI-12 network is is connected together to a data logger. controlled by the data recorder.

5.3 PC Data Acquisition System Q-Log Application 2 3 4 05404 -03539 04085 -033 (3 Part No. USB-RS485-Pro legilated RS-485 to USB m Part No VibWire-108-485 8 channel Vibrating Wire Se RS-485 digital network.

Additional Information

The USB-SDI12 media converter only controls the network lines. Individual instruments use their own local power supply and are active continuously.

Figure 5 below shows how a typical multi-instrument data logger system is wired together.

Multi-instrument Data Logger Solution

As soon as a measurement instruction is sent by the data logger the the SDI-12 bus is activated and all the instruments are







RS-485 Network Connection

6.1 Earth Connection

Ensure that a good Earth connection is made and fitted to each instrument in order that the lightning protection discharge tubes will operate.

Lightning protection is provided for all VW sensor inputs and between the RS-485 - 0V pins.

Important Note

Failure to fit the Earth connection will prevent the lightning protection from operating.

6.2 PC Data Acquisition System - RS-485 Solution

Additional Information

A PC based RS-485 system can operate with up to 32 instruments and to a direct network span of up to 1000 m. This is an ideal solution for large distributed applications. Use an external power supply for more 3 units on a network string.

The USB-RS-485 media converter only controls the network lines. Individual instruments use their own local power supply and are active continually.

O-Log Application Image: State of the state



12 - 16 V DC Connect external power supply if the maximum current for the number of devices on the RS485



For applications requiring a large number of distributed instruments then simply adding a new USB-RS485 media converter to a PC will enable multiple network systems to operate together.

The Keynes Controls Q-LOG software enables unlimited expansion by adding new networks to a PC at any time. In fact the software is flexible enough to enable SDI-12 and 485 networks to coexist enabling many different intelligent sensors to be used.

7.0 Modbus RS-485 Network Connections

The VibWire-108-Modbus connects onto the RS-485 network in exactly the same way as using the standard RS-485 network for communications

Fig 8



8.0 PC/Laptop Data Recording & Display System



The image above shows how the VibWire-108 can be integrated into a PC based data recording and display solution using the Q-LOG application software. The VibWire-108 will operate with 3rd party software and hardware supporting the correct network protocols.

The USB-SDI12-Pro and USB-RS485-Pro media converters are fully integrated into the Q-LOG application software, and can also be used to power the vibrating wire sensor interfaces directly from the PC USB port.

The Q-LOG software will also operate with many third party intelligent sensors .

0 - 5 Km depending upon signal cable quality

SDI12 Data

12 DC

9.0 Technical Specifications

The technical specification for the different models is shown below.

All of the VibWire-108 family of products use the same terminal port setting for configuration operations.

Measurement Data:		
Number of channels	8 x 4 Wire VW Inputs - User Selectable	
VW sensor coil resistance	to 2 K Ohm (standard):- other ranges on request	
Distance of VW sensor to interface	0 10 Km depending on cabling.	
Frequency range	400 - 6 KHz (standard) Other ranges on request	
Frequency Resolution Accuracy	32 bit resolution 0.001 Hz	
Long term stability	± 0.05 % FS max / year	
Temperature range	- 50 to 70 Deg C	
Temperature resolution	0.1 °C +/- 0.2 Deg Thermistor 10 K Ohm standard 3.3 K Ohm on request	
Temperature accuracy	± 0.2 °C / 0.2 °F SDI-12	
Thermistor measurement	A half bridge ratio-metric measurement. Value returned in mV. Is used for temperature compensation on VW measurements using Steinhart-Hart thermistor equation or beta value.	
Thermistor excitation	2.5 V DC 50 ppm /Deg C	
Input resistance	10 K Ohm 0.1 % Completion resistor (Standard) 3.3 K Ohm on request	
Units	Freq (Hz), Digits (Hz ²), SI Units, Temperature Deg C, mV	
Display only - Resolution	5 digit - 0.1 Hz	
Electrical Data		
Voltage supply	SDI-12 10.5 to 16V DC	
Current compensation SDI-12 Option only	Typical values are @ 12 V DC Excitation	
Idle mode Active / measurement	1.2 mA 8 mA data transmission 58 mA including frequency display These values may change slightly between sensors. Use figures as a guide only.	
Measuring time warm up response	500 ms 3 seconds per channel depending on the VW sensor being used (Typical)	
Length of data lines SDI-12	0 100 m	
SDI-12 Address mode	Supports enhanced addressing 0 9 A Z	
Dimensions (mm)	L_260 W 127 D 38	
Material		
SDI-12 Digital Port	SDI-12, 1200 Baud, 7 bit, N stop bit, Even Parity - other speeds on request	
CE Conformity	CE conformity according to EN 61000-6	
Weight	400 g	
Communications		
Terminal Port	9 Way Male - 9600 Baud 8 data, No Parity, 1 stop bit, No Flow control - DTE	
SDI-12 Digital Port	1200 Baud, 7 bit, N stop bit, Even Parity - other speeds on request	
RS-485 Network Settings	1200 Baud, 7 data bit, N stop bit, even parity	
RS-485 Network Settings - modbus	9600 Baud, 8 data bit, 1 stop bit, even parity	

10.0 System Test - Power on sequence

The instructions are the same for all models.

- Step 1 Power on the VibWire-108.
ResponseApply 12V DC to the power connector shown on Fig 1, Page 6
The VibWire-108 will briefly show the message 'HELLO' on the 5 Digit LED Display.
- Step 2 The display will default to '0' on the LED display.

The instrument will wait until a start measurement command is received before a measurement is made.

11.0 VibWire-108 Digital Communications

The instructions below detail the operations to follow to operate the VibWire-108 across both the SDI-12 and RS-485 serial networks.

No sensor configuration details need be applied to the VibWire-108 when operating with the cable free transmitter, RS-485 or SDI-12 network. Simply connect the sensors to the interface as shown in Fig 2 and initiate the commands listed below.

Recommended Test

Use a single instrument only when undertaking initial measurements with a VibWire-108 on the RS-485 or SDI-12 network. This simplifies the software and will speed up the understanding of the command used to obtain data. It is very easy to test the results measured across the RS-485 and SDI-12 network with the ones shown on-board frequency display of the unit.

The results obtained across the RS-485 and SDI-12 network will be same as those shown on the display for a specified channel.

The default instrument address for a unit straight out of the box is 0 - results will be random numbers with no sensors installed.

Test Measurement

Issue command **OM!** to start measurement operations. The VibWire-108 will scan all channels **OD0!** returns items of data 0+ Freq Chan 0 + Freq Chan 1 + Freq Chan 2 + Freq Chan 3

Ensure that each instrument used on a network has a unique ID number assigned within its configuration in order to correctly identify the data that is being recorded.

11.1 Command Structure & Operations across an RS-485 network

The VibWire-108 uses a command structure across the RS-485 network very similar to that used by the SDI-12 network in order to simplify the overall system operations. Understanding the control of the instrument on one network be that RS-485 or SDI-12 will make using the the unit on the other a very simple operation.

The RS-485 and SDI-12 network both operate at the same speed of 1200 baud. The RS-485 network speed can be increased at the time of manufacture. Currently it is factory set.

Even though this is a relatively slow rate, as networks go, it is more than adequate for the small amounts of data transmitted by the instrument. No break characters are transmitted in an RS-485 command and any sent will be ignored by the instrument.

A short delay of approximately **10 ms** is added between a command received by VibWire-108 and its transmission of data since this delay is used to allow time for the host PC to turn off its transmitter when using soft-negotiation for data flow control operations. Under normal RS-485 data transmission operations the RTS line on the serial port is used for flow control operations.

Keynes Controls recommends an RS-485 interface with hardware negotiation is used to control the VibWire-108 across a network.

11.2 Timing Constraints RS-485 / SDI-12 Network

There are no timing constraints for the transmission of instructions and receipt of data across an RS-485 network compared to the operations on and SDI-12 network.



Figure 10 shows timing delay used on SDI-12 serial bus prior to sending data

Typically the VibWire-108 takes 5 seconds to complete the scan of the first sensor and a further 3 seconds for any other sensor connected to the instrument. The actual response time for the instrument is dependent upon number of sensor fitted and can interrogated using the aM! command.

11.3 RS-485/ SDI-12 Commands

The commands used by instruments on the SDI-12 and RS485 network are the same

In the following commands 'a' and 'b' are the address of the instrument and can only be integers 0 to 9 or the characters a - z.

where

Data Access Time

'ttt' represents a time in seconds (0 to 999 seconds)
'n' or 'nn' represents a number of channels (00 to 99 channels)
\r and \n are the Carriage Return and Line Feed characters - ASCII 13 and 10.

11.4 Assigning Data Transmission over the SDI-12 or RS-485 Network

All of the VibWire-108 models use the **SERAL** option to assign the data transmission operations across the digital network. A 10 minute time-out feature ensures that the instruments cannot be left displaying real-time frequency results.

For Modbus operations the instrument scans automatically at the pre-set sample time as soon as power is applied, see page 24 for setting the sample time.

The Modbus ID is set exactly the same as for the SDI-12 and normal RS-485 operations.

11.5 Starting Data Acquisition Operation on a SDI-12 Network

To activate the analogue output channels on the VibWire-108.

1. Starting at



Instrument Identifier

upon request.

Each instrument deployed on the SDI-12 multi-drop network must have a unique instrument identifier set in order to identify specific instrument on the network. For the SDI-12 network this identifier is 0-9. The SDI-12 ID can be expanded upon request.

See Page 8 for details on setting the ID number.

- 2. Select "Menu In" button
- 3. Use the Up & Down Keys to select the option "SERAL"

"Analg C0d C1d C2d C3d C4d C5d C6d C7d" are the other options available

Once the "Seral" output is selected the "Menu Out" key has to be pressed to confirm this option.

4. The VW-108 will return to the display



must have a unique instrument identifier (ID) set in order to

identify specific instrument on the network. For the RS-485 network this identifier is 0-31. The SDI-12 ID can be expanded

See Page 26 for details on setting the ID number using the

keyboard menu system. System Maintenance option of the



and now the SDI-12/RS-485 port for the instrument is now activated.

11.6 Starting Data Acquisition Operation on a 485 Network Instrument Identifier

To activate the SDI-12/485 output channels on the VibWire-108 Each instrument deployed on the RS-485 multi-drop network

1. Starting at



- 2. Select "Menu In" button
- terminal menu system page 23.
 Use the Up & Down Keys to select the option "SErAL"

"Analg C0d C1d C2d C3d C4d C5d C6d C7d" are the other options available

Once the "SErAL" output is selected the "Menu Out" key has to be pressed to confirm this option.

4. The VW-108 will return to the display



and now the SDI-12/RS-485 port of the instrument are now activated.

11.7 Modbus version instrument

The Modbus version instrument will start the scan sequence automatically as soon as power is applied. The scan period is set using the built in menu system using the keyboard. See page 24, section 23.3 for details.

12.0 Instrument Identifier

Each instrument deployed on the multi-drop network must have a unique instrument identifier set in order to identify specific instrument on the network:

For the RS-485 network this identifier is in the range 0-9 / a-z. For the SDI-12 network then the ID number is in the range 0..9 - Additional ID numbers are supported a .. z. For Modbus operations the ID number is currently limited to 1 .. 32.

13.0 Start Measurement Commands

There are 2 separate commands supported by the VibWire-108 for initiating measurements across an RS-485 network and are named '**aM**!' and '**aC**!'. Table 1 includes a complete description of the commands used by the VibWire-108

The '**aM**!' starts a measurement and responds as soon as the data is ready to be transmitted from the instrument. This command returns all instrument sensor inputs as a string

The '**aC!**' command starts concurrent operations that are used to initiate measurements upon multiple instruments deployed across the network. The 'aC!' command frees the network bus so that other devices can operate freely.

14.0 Advice on the choice of measurement commands

The VibWire-108 supports both the individual and concurrent start measurement commands.

Keynes recommends using individual start measurement commands where there are large distances between devices and the network cable installation quality is poor. Should there be substantial voltage losses on the supply cable then the additional load of a lot of sensors scanning simultaneously may cause errors with some instruments being unable to operate correctly.

For fast results and small scale systems then the concurrent start measurement command can be used.

14.1 Possible Network Problems

Should a larger than expect load be put onto the SDI-12 network then the voltage drop between 0 V and SDI-12 12 V supply can be as high as 2 V. If this is the case then the instrument may not respond as expected.

15.0 SDI-12 Supported Commands

The following commands are supported by the VibWire-108 SDI-12 model

Description Acknowledge active

Send ID:

provided to complement SDI-12 protocol
Address Query

identifies instrument address and commonly used on single instrument operations only.

Change Address:

used to change instrument address from **a** (inital) to **b** new ID for network operations

Start Measurement

instruct an instrument to make measurement

Concurrent measurement:

Used for starting a measurement for all instruments on a network at the same time.

This command frees RS-485 bus for other devices

Send data

data returned aND! = Vib + Vib + Therm + Therm and has same format for each command

Thermistor 1 & 2

Thermistor Type 1 Temperature sensor settings

Parameters from the sensor calibration sheet

Steinhart-Hart Parameters Thermistor resistance/temp calculation

Thermistor Type 2 Temperature sensor settings

Parameters from the sensor calibration sheet

Steinhart-Hart Parameters Thermistor resistance/temp calculation

Page 36 shows sample calibration data sheet

VW Sensor Input Channel Settings

Therrmistor Type

The VW108 supports 2 different thermistor types for temperature measurement.

Thermistor Temperature Calculation

Master

a! al!

?!

Used to make command set SDI-12 compatible

aAb!

a = initial address **b** = new address

aM!

a = address of instrument example 0M! starts scan for ID 0

aC!

start measurement instrument address a

aD0! aD1! aD2! or aD3!

 $aD0! = channel 0 and 3 \quad VibWire Sens \\ aD1! = channel 4 and 7 \quad VibWire Sens \\ aD2! = channel 0 and 3 \quad Therm/analog \\ aD3! = channel 4 and 7 \quad Therm/analog \\$

VibWire-108 supports 2 thermistor types

aXT1RE! aXT1T0! = 25 aXT1BET!

aXT1ST0! aXT1ST1! aXT1ST2! aXT1ST3!

aXT2RE! aXT2T0! = 25 aXT2BET!

aXT2ST0! aXT2ST1! aXT2ST2! aXT2ST3!

aXCH0FN!

F = Frequency typeN = VW Channel 0 .. 7

aXCH0TN! = Thermistor type

where a = ID T = Thermistor Type N = Thermistor Channel Input = 0..7

aXT1TYn!

a = ID n = integer 0 .. 2

VibWire-108 Response

a∖r∖n

a13KEYNESCOVibWire-1080001\r\n

Part Description assigned by Keynes

a∖r∖n

Where a = ID number 0 - 9 (standard) / (a..z) Enhanced SDI-12 0 - 9 / a - z for RS485

b\r∖n

 $\mathbf{a:b}$ = number 0 - 9 or a - z

a0268\r\n

instrument with address a returns 8 x vibwire & 8 x temp after 60 seconds

a0268\r\n

initial response only after receipt of instruct and no response when data ready to be sent.

+xxxx.x+xxxx.x+xxxx.x+xxxx.x\r\n

Resistance at 25 Deg C T0 - generally 25 Deg C Beta Value

A in Steinhart-Hart B in Steinhart-Hart C in Steinhart-Hart D in Steinhart-Hart

Resistance at 25 Deg C T0 - generally 25 Deg C Beta Value

A in Steinhart-Hart B in Steinhart-Hart C in Steinhart-Hart D in Steinhart-Hart

- 0 = output in Hz
- 1 =output in digits = F^2/1000
- 2 = use formula A + B*digits + C*digits^2 + D*temperature

digits = Frequency² in units of Hz²

0 = Voltage ratio 1 = Type 1 thermistor (use XT1RE etc. as above) 2 = Type 2 thermistor 11 = Type 1 resistance ratio, output Rt/R25 12 = Type 2 resistance ratio, output Rt/R25 99 = Output mV at terminal 0 = resistance ratio - thermistor data sheet (Rt/R25)

- $\begin{array}{l} 1 \ = \ Beta \ value \ calculation \\ 1/T \ = \ 1/T_0 \ \ + \ log(r)/Beta \end{array} \qquad where \ r \ = \ R_t/R_{25} \end{array}$
- 2 = Steinhart-hart equation

 $1/T = A + B(Ln R_t/R_{25}) + C(Ln R_t/R_{25})^2 + D(Ln R_t/R_{25})^3$

16.0 RS-485 Supported Commands

The instrument commands for the RS-485 and SDI-12 versions of the instrument are identical except for the "%" prefix at the start of the command. See table below

Description	Master	VibWire-108 Response
Acknowledge active	%a!	a\r\n
Send ID:	%al!	a13KEYNESCOVibWire-1080001\r\n
provided to complement SDI-12 protocol		Part Description assigned by Keynes
Address query	%?!	a\r\n
identifies instrument address and commonly used on single instrument operations only.	Used to make command set SDI-12 compatible	Where a = number 0 - 9 for SDI-12 0 -9 letters a - z for RS485 A - Z
Change Address:	%aAb!	b\r\n
used to change instrument address from ${\bm a}$ (inital) to ${\bm b}$ new ID for network operations	a = initial address b = new address	a : b = number 0 - 9 or a - z
Start Measurement	%aM!	a0268\r\n
instruct an instrument to make measurement	 a = address of instrument example 0M! starts scan for ID 0 	instrument with address a returns 8 x vibwire & 8 x temp after 60 seconds
Concurrent measurement:	%aC!	a0268\r\n
Used for starting a measurement for all instruments on a network at the same time.	start measurement instrument address a	initial response only after receipt of instruct and no response when data ready to be sent.
This command frees RS-485 bus for other devices		
Send data	%aD0! aD1! aD2! or aD3!	+xxxx.x+xxxx.x+xxxx.x+xxxx.x\r\n
data returned aND! = Vib + Vib + Therm + Therm and has same format for each command	aD0! = channel 0 and 3 VibWire Sens aD1! = channel 4 and 7 VibWire Sens aD2! = channel 0 and 3 Therm/analog aD3! = channel 4 and 7 Therm/analog	
Thermistor 1 & 2	VibWire-108 supports 2 thermistor types	
Thermistor Type 1 Temperature sensor settings	%aXT1RE!	Resistance at 25 Deg C
Parameters from the sensor calibration sheet	%aXT1T0! = 25 %aXT1BET!	T0 - generally 25 Deg C Beta Value
Steinhart-Hart Parameters Thermistor resistance/temp calculation	%aXT1ST0! %aXT1ST1! %aXT1ST2!	A in Steinhart-Hart B in Steinhart-Hart C in Steinhart-Hart
See page 36	%aXT1ST3!	D in Steinhart-Hart
Thermistor Type 2 Temperature sensor settings	%aXT2RE!	Resistance at 25 Deg C
Parameters from the sensor calibration sheet	%aXT2BET!	Beta Value
Steinhart-Hart Parameters Thermistor resistance/temp calculation	%aXT2ST0! %aXT2ST1!	A in Steinhart-Hart B in Steinhart-Hart C is Steinhart-Ilart
See page 36	%aXT2ST3!	D in Steinhart-Hart
VW Sensor Input Channel Settings	%aXCH0FN!	0 = output in Hz
	F = Frequency type N = VW Channel 0 7	2 = use formula A + B*digits + C*digits^2 + D*temperature
		digits = Frequency ² in units of Hz ²
Therrmistor Type	%aXCH0TN! = Thermistor type	0 = Voltage ratio
The VW108 supports 2 different thermistor types for temperature measurement.	where a = ID T = Thermistor Type N = Thermistor Channel Input = 07	1 = Type 1 thermistor (use XT1RE etc. as above) 2 = Type 2 thermistor 11 = Type 1 resistance ratio, output Rt/R25 12 = Type 2 resistance ratio, output Rt/R25 99 = Output mV at terminal
Thermistor Temperature Calculation	%aXT1TYn!	0 = resistance ratio - thermistor data sheet $(R_{\rm I}/R_{\rm 25})$
	a = ID n = integer 0 2	$ \begin{array}{l} 1 = Beta \ value \ calculation \\ 1/T = 1/T_0 \ + \ log(r)/Beta \end{array} \qquad where \ r = R_l/R_{25} \end{array} $
		2 = Steinhart-hart equation
		1/T = A + B(Ln Rt/R25) + C(Ln Rt/R25) ² + D(Ln Rt/R25) ³

17.0 Examples Of Using RS-485/SDI-12 Commands

The following examples show how to undertake the various tasks needed to setup and make readings across the RS-485 and SDI-12 networks. The command structure between the SDI-12 and RS485 models is essentially the same except all RS-485 commands use the '%' sigh at the start of all instructions.

The SDI-12 networks only supports up to 10 instruments with address range: 0 to 9 unless otherwise stated.

17.1 Changing the ID Number (address)

The following example demonstrates how to change the instrument ID number from the default factory setting 0 to 5.

Use the command 'aAb)!'	where a = Start ID	b = Final ID
SDI-12 master sends:	'0A5!'	Instrument responds	 5\r\n Return New Line (5 representing new ID number 5\r\n Return New Line (5 representing new ID number
RS-485 master sends	'%0A5!	Instrument responds	

17.2 ID Number Query

This command has been included to remain compatible with the SDI-12 and should be used for used with single instrument operations only. Useful command when identifying ID numbers for instruments to be deployed on a multi-instrument network.

The example below is to show the ID number of a single instrument

Use the command '?!' . The '?!' command only works when a single instrument is in operation.

master sends: '?!' Instrument responds 3\r\n Return New Line (3 is the ID number)

17.3 Start measurements for Instruments on a network

The following example shows how to start measurements on instruments with ID numbers 2, 7, and 9 respectively.

For this example the instruments are instructed to start readings one at a time and the network is not freed up until each instrument responds that the readings are being undertaken.

The instruments will start their measurement operations but will not send data across the network until instructed to do so.

Use the command ' aM! '	where a = Instrument ID Number
Use the command '%aM!'	for RS-485 network operation

Examples of use.

The following example is based upon a simple application of 3 x VibWire-108 units connected together on a local SDI-12 network. Unit 1 with address 2 has 4 vibrating wire sensors, Unit 2 with address 7 has 6 sensors connected and finally Unit 3 has only 2 sensors connected.

master sends:	'2M!'	Instrument responds followed by	'20144\r\n' '2\r∖n'	indicated readings available after 60 secs when the measurement is completed
	7M!		'70206\r\n' '7\r\n'	
	9M!	، ۱	90082\r\n' 9\r\n'	

Note. For this command the RS-485 network will not become available until each instrument completes its measurement cycle.

The VibWire-108 is fully integrated into the Keynes Controls AquaLOG SDI-12 data recorder and communications interface.

The menu opposite shows a typical SDI-12 network command as used by the AquaLOG to start a measurement on an instrument with ID=4 and used to download 8 x frequency inputs and 8 x temperature levels.

A power-up delay of 500 ms is needed before a measurement can be made.

SDI12 Setup

1 Powerup delay (ms) 2 Powerdown delay (ms)	500 0		
3 Device 1 Command	[D] 4M!	4D0!	4D1! 4D2! 4D3!
4 Device 2 Command			
5 Device 3 Command			
6 Device 4 Command			
7 Device 5 Command			
8 Device 6 Command			
9 Device 7 Command			
A Device 8 Command			
B Device 9 Command			
C Device 10 Command			
U Up. Т Тор			

17.4 Start measurements using the concurrent command

The VibWire-108 supports the aM! and aC! measurement commands. The concurrent measurement '**aC**!' command differs from the '**aM**!' command as it frees the network after the initial command response to allow other devices to operate. Concurrent measurements enable multiple instruments to respond faster to measurement commands on a network. The disadvantage of using the '**aC**!' command is that there can be greater power loss on the power supply network lines, so good quality network cabling for long network strings should be used.

The '**aC**!' command initiates the measurement cycle within the instrument to start reading from the sensors however the data still has to be requested from the VibWire-108 before being sent across the network.

Example of concurrent measurements for instruments with ID numbers 1, 6, and 7 respectively.

For this example the instruments are instructed to start readings one at a time and the network is not freed up until each instrument responds that the readings are being undertaken.

The instruments will start their measurement operations as soon as the command is received but will not send data across the network until instructed to do so.

Use the command ' aC! '	where a = Instrument ID Number.	
master sends: '1C!' - 4 sensors	Instrument responds '1 0144\r\n ' indicated readings available after 14 secs The network is free for other devices as soon as this response is returned.	
'6C!' -3 sensors	'6 0113\r\n '	
' 7C!' - 5 sensors	'7 0175∖r∖n '	

17.5 Read data values from the VibWire-108

No matter which instruction 'aM!' or 'aC!' is used to initiate measurement operations for the VibWire-108 has to be instructed to send data when it becomes available. It takes the instrument 60 seconds to make sensor values available after being instructed to make a measurement. The vibrating wire readings are in Units Hz. The Temp/Current loop input are in Units mV..

Use the command:	'aD0!'	 Vibrating Wire inputs 0 - 3
	'aD1!'	 Vibrating Wire inputs 4 - 7
	'aD2!'	 Temp/current loop inputs 0 - 3 (values in mV)
	'aD3!'	 Temp/current loop inputs 4 - 7 (values in mV

Instrument responds: 'a+xxxx.x+xxxx.x+xxxx.x+xxxx.x/r/n' xxxx.x is the format of the number returned - 1 decimal place

for example to read all the sensor data back from an instrument with ID = 4

master sends: '4D0!'	Instrument responds: '4+1011.3+1204.4+1101.3+1190.7'	Vibrating wire data
'4D1!'	Instrument responds: '4+1021.5+0000.0+1141.2+0000.0'	0000.0 is returned when no sense
installed	-	

installed

17.6 Temperature/Current loop Data Format

For an instrument with 7 VW sensors installed.

'4D2!' Instrument responds: '4+0050.6+0056.1+0101.2+0017.0' shows results with only 7 temperature values mV '4D3!' Instrument responds: '4+0051.4+0058.3+0110.2+0000.0'

No Data is available Instrument responds 'a\r\n' or this example '4\r\n'

Note. The temperature values are in mV only. Thermistor linearisation is needed is convert the results into engineering values.

18.0 Connection to an Analogue Data Acquisition System

The following details show how to configure and optimise the VibWire-108 analogue outputs to operate with a analogue input data acquisition system or logger unit.

Technical Specifications - Analogue Output Ports

- 8 x 0 2.5V DC single analogue output ports 16 bit DAC
- 8 x thermistor outputs 3.3 KOhm completion resistors

User Control Keys

Fia 11

Menu-out

Down

Up

000000000000000

Menu-in



The VW-108 can be connected to an external data acquisition system or data logger using the analogue output ports fitted onto the instrument. In order that the correct values can be interpreted by the logger/acquisition system they are first scaled into a suitable analogue signal by the VW-108 before being passed on for measurement. Each output channel can be uniquely configured to support any manufactures sensor.

When defining the operation of the analogue output each channel has to have the sensor operating characteristics defined. For the VW-108 this means that the minimum operating frequency and span are set into the instrument.

Once the operating frequencies for the sensor are assigned the instrument scales the measured sensor frequency over the range 0V = minimum frequency and 2.5V = maximum frequency.

18.2 Connection to an Analogue Input or Data Acquisition System

The analogue output ports are singles ended and as such, care should be taken when connecting to a differential input channel. The example shown in Fig 9 shows the analogue output from the VibWire-108 connected to the AquaDAT analogue data

- Sense = 0V (single ended) or -Vin (Differential Input) + Sense = +Vin

18.3 VibWire-108 Analogue Port Configuration

Low Frequency := 500 - 3000Hz defined in 100 Hz intervals := 100 Hz steps. Range

18.4 Starting Analogue Output Ports

To activate the analogue output channels on the VibWire-108

1. Starting at



- 2. Select "Menu In" button
- 3. Use the Up & Down Keys to select the option "Analg"

"Serial C0d C1d C2d C3d C4d C5d C6d C7d" are the other options available

Once the "Analg" output is selected the "Menu Out" key has to be pressed to confirm this option.

The VW-108 will return to the display 4.



and now the analogue output channels for the instrument is now activated.

Each of the vibrating wire sensor inputs can be individually configured. Setting the analogue output channel is only needed when using the instrument with an external data logger or analogue acquisition system and is not required when measurements are to be made across the SDI-12 and RS485 networks.



acquisition unit. The AquaDAT supports a full differential or

single ended input measurements. The device also features an

auto-ranging inputs to optimise measurements.

AquaDAT



Analogue output on VW-108

Figure 9

Analogue Data Acquisition

0V Vin

+Vin +2.5V

0V - Vin +Vin +2.5V

0V - Vin +Vin

+2.5

18.5 Optimising the Analogue Output Settings

Example 1

The VibWire-108 contains 8 independently configurable analogue output ports and they are used to represent the output signal from the sensor.

Each analogue output is of the range 0 - 2.5V DC and any analogue output must scale a result to within this range Care should be taken to ensure that the output signal is scaled as close as possible to sensor range

For example, Channel 0 is used to output a signal from a sensor with operating range of 1452 - 3176 Hz

It is not possible to set the output range of the DAC directly to represent the absolute range of the sensor and so it must be set

to cover the sensor range with the minimum overlap in order to obtain the highest resolution.

a range of

0V = 1400 Hz & 2.5V = 3200 Hz so CH0 LF = 1400 and CH0 RA = 3200 - 1400 = 1800 Hz

will give the highest resolution for this example

DAC Resolution output port = 16 Bit so Frequency Resolution = 1800 / 65536 = 0.03 Hz

in practice accuracy of around 0.5 Hz can be achieved when connecting the VW-108 to an analogue data acquisition system after allowing for the losses due to the Digital-analogue and Analogue-digital conversion process. The 0.5 Hz measurement accuracy is achieved using the AquaDAT as the input signal measuring device.

Only when operating the VibWire-108 with an active analogue output port need the operating characteristics for the vibrating wire sensor be defined.

For general purpose operations the analogue output should be set to represent the full operating range of the sensor.

18.6 Integration to the AquaDAT Analogue Sensor Interface

Example 2

A vibrating wire pressure sensor with operating frequency 400 Hz to 1000 Hz connected to channel 5 on the VW-108 and the analogue output is to be connected to an AquaDAT Sensor interface.

CH5 LF = 400 CH5 RA = 600 (where range = 1000 - 400) and CH(0-7).RA is the range parameter.

the AquaDAT input channel range is to be set to 2.5 V

therefore 0V = 400 Hz and 2.5V = 1000 Hz

The AquaLOG will auto-range to optimise the signal measurement

The data logger will scale the results over the full range Resolution = 600/65536 = 0.01 Hz

In practice an measurement accuracy of 0.05 Hz will be achieved after allowing for losses in the analogue conversion process.

Unit Conversions

Celsius to Fahrenheit	(°C × 9/5) + 32 = °F
Fahrenheit to Celsius	(°F - 32) x 5/9 = ℃

Example: Convert 26° Celsius (a nice warm day) to Fahrenheit

First:	26°×9/5	=234/5	=46.8
1 11 01.		-20-4/0	-+0.0

Then: 46.8 + 32 = 78.8° F

19.0 Real-time Frequency Display

All of the VibWire-108 models contain a 5 digit 7 segment display and this can be used to display the instantaneous frequency from any of the individual vibrating wire sensor inputs.

Vibrating wire sensors can be deployed a considerable distance from the VibWire-108 interface and may well be embedded into a structure. To ensure that the sensors are operating correctly, simply observe the sensor operating frequency on the 7 segment display and then confirm the result is within the operating range as specified by the manufacturer.

When operating in a real-time mode the instrument frequency display responds instantly to effects upon the sensor.

To use the VibWire-108 as a real-time frequency display follow the instructions below:

19.1 Assigning the real-time display to a VW sensor input

To activate the real-time frequency display

1. Starting at



The "Basic" menu item is the first menu item available after the instrument is powered on.

- 2. Select "Menu In" button
- 3. Use the *Up* & *Down* Keys to select the option



The Display above shows the option required to place Channel 0 for real-time frequency output

the other options available are:

"Analg Seral C0d C1d C2d C3d C4d C5d C6d C7d" C0d = Channel 0 C7d = Channel 7

Once the "Cod" option is selected then the "Menu Out" key has to be pressed to confirm this option.

4. The VW-108 will now display the real-time sensor frequency for channel 0.



The above example shows a typical real-time frequency result .

20.0 Loud Speaker



All of the VW-108 range of instruments are supplied with an internal ceramic speaker. The speaker can be activated and the sensor ping and resultant echo can be heard.

The speaker used in collaboration of the frequency display should enable nearly all sensors to be tested no matter their location using only the VW-108 interface unit.

Sensor problems

Should a clean ping not be heard when the vibrating wire sensor is being sampled by the instrument the following guide should help.

- 1) If there is only random noise on the speaker for the defined channel then check the wiring and circuit resistance. The most common error is an open circuit. Locate and fix the broken cable.
- 2) If a ping can be heard but it is faint then the sensor cable may be too long, or a to high cable resistance is being used causing degradation of the signal amplitude. Finally the gauge sensitivity may be to low.
- 3) If the ping is not a pure tone then the gauge is possibly faulty. The gauge may have become damaged during installation.
- 4) If a low frequency hum is heard then noise pickup can be a problem. If the gauge cabling is routed near a transformer, electric motor, high current power cables, etc, then relocate or reorient the gauge for minimum pickup. Ensure that only shielded cable is used and that the shielding is terminated at a single point to prevent capacitive pickup.

21.0 Vibrating Wire Sensor Installation

The vibrating wire sensors are connected directly into the VW Sensor Input channels as shown below. The instrument contains a completion resistor for the thermistor sensor enabling the temperature reading to be made along with the vibrating wire sensor readings. The VibWire-108 can be used with most different thermistors used within the vibrating wire sensors.

Connection to the instrument is as follows:





In order to ensure that there are sufficient points to terminate sensor sheathing when armoured cable is used to connect a sensor to the VibWire-108 the following terminal points are internally wired in common:

Earth Earth Earth Earth Gnd

On the power supply connection terminals

Therm -Sense -

on both the sensor input and analogue output terminals.

Fig 14A shows how the channels are wired together to form common earth connection. Effectively all of the Earth, Gnd, S_{ense} - and T_{herm} - terminals are wired together.

Any earth Sheathing from armoured cable etc.. can be connected to any of terminals mentioned above for ease of installation.

21.4 Lightning Protection

The lightning protection within the VibWire-108 cannot protect the instrument from a direct lightening strike. It is used to protect the instrument from local ground strikes close to the sensors and cabling.

All of the sensor inputs are protected by transorb and gas discharge tubes. The transorb are high capacitance devices and are not used on all systems as they can distort low level signals to a point where the instrument can not be accurately measured. The transorb does protect the instrument at lower levels than the gas discharge tube, and starts to become active around 12V.

The gas discharge tube protection activates at around 92V DC and resets instantaneously after the lightning strike effect has died away.

Fig 14 shows the VibWire-108 connected to a system earth using the Earth terminators mounted adjacent to the power connectors.



21.2 PCB Jumper Settings



All of the VibWire-108 models support both thermistor and analogue inputs for the range 0 - 2.5 V DC and current loop inputs.

For current loop operations such as those 0- 20 mA, 4- 20 mA then external excitation is required.

Jumper Open	=	0-2.5V DC
		4-20 mA loop
Jumper Closed	=	Thermistor

21.3 Multiple Instrument Installation



Figure 17 shows how to Earth multiple instruments within a single enclosure. Ensure that good quality cable of around 2.5 mm diameter core is used to connect the instruments earths together and that a good connection to a main system earth is obtained. The Earth connection is essential for the lightening protection to work.



Fig 18



9 Pin - DTE

것것것것 같겠겠겠 것것것것

Models VibWire-108-SDI12, VibWire-108-RS485, and VibWire-108-Modbus can be configured using the instrument terminal port.

The following instructions are for the Microsoft Windows Operating system.

Step 1:

Connect the PC/Laptop to the VibWire-108 using the USB-RS232 interface and null modem cable as shown above. The terminal port is configured as a 9-way DTE device.

Step 2:

Plug the USB-RS232 adapter into the PC/Laptop.

From the operating system control panel select the "device manager" option. A Window similar to that shown opposite will appear.

Select the 'Ports (COM & LPT)' option from the menu list to identify the Comm port number used by the USB-RS232 interface.

Port Comm in use by the USB-RS232 media converter

Connect To	COMI Properties Pot Settings Pot Settings Rts per second: 9600 Rts per second: 9600 Refty: Eleme Refty: Ele
Cognect using: COM1 OK Cancel Microsoft Hyper-terminal - Connect To Window	Bow control: None Bestore Defaults OK Cancel Apply Properties Window

The Terminal Port Menu System

There is no software required with this device part The menu system can be accessed and used by any modern terminal emulator software such as Microsoft Hyper-terminal or Token-2 etc. The terminal software has to be VT100 compatible to operate correctly.

> The example Windows above, are taken from the Hyper-terminal software, however the communication port settings are the same no matter which package is used.

Step 3

Start the Terminal emulator software and configure the communications port to 9600 Baud, 8 data bits, 1 stop bit, No parity.

The Comm port number used by the USB-RS232 media converter is shown in the Windows 'Device Manager' Window.



Microsoft Windows 7 Device Manager Window

22.1 **Terminal Port Operation**

The terminal port built-in to the VibWire-108 enables the instrument to be easily configured using the built in menu system to set all the calibration parameters. 22.2

from a terminal emulator package, which is frequently a feature included most operating systems.

Each VW sensor input channel can be individually configured using details taken directly from a sensor calibration data sheet.

Currently the VibWire-108 supports 2 pre-defined temperature sensors configuration options that can be assigned to each VW sensor input.

Step 4

Power on the VibWire-108, see instructions on page 6 with the terminal port connected to a laptop as shown on page 22.

After the 'HELLO' message flashes on the 7 segment display and the instrument is connect to the PC/Laptop as shown on page 22, then the main menu will appear. The Comm port setting are shown in the specification table.

Main Menu

1 System Maintenance 2 Thermistor type 1 3 Thermistor type 2 4 Diagnostics 5 Channel 0 6 Channel 1 7 Channel 2 8 Channel 3 9 Channel 4 A Channel 5 B Channel 5 B Channel 6 C Channel 7 U Up. T Top.

Fig 20

22.3 Selecting the VW sensor output units.

The following procedure is for the VibWire-108-SDI12, VibWire-108-RS485, and VibWire-108-Modbus models only.

The VibWire-108 supports output units of Hz, Digits (Hz²), and SI units.



Important Note

The 'Thermistor Type' calculation option determines which calculation is used regardless to the calibration parameters that are configured.

22.4 Temperature Sensor Configuration

The VibWire-108 uses the in-built thermistor inside a VW sensor to measure temperature. The menu system enables two different thermistor sensor details to be predefined and stored into the instrument.

.The menu system enables both the VW sensor linearisation and thermistor temperature equations to be configured. The thermistor resistance and therefore temperature is calculated using the Steinhart-Hart equation. The output SI unit is in Deg C.

The VW sensor linearisation uses the industry standard quadratic equation to convert the VW signal into engineering units.

The parameters for this part of the instrument configuration can be found on the VW sensor calibration data sheet.

For lower accuracy temperature readings or when the calibration factors are not known then the thermistor Beta value, T_0 and R_0 parameters can be assigned.

Common VW Sensor Thermistor Part Numbers

YSI 44005 Vishav 1C 3001 B3	The part numbers are for 3 K Ohm thermistor commonly used by most different VW sensor manufacturers to measure temperature
RS Part no: 151-215	The sensors give 3 K Ohm resistance at 25 Deg C

The most common material used in these sensors uses material type F from GE sensing.

Beta Value temperature

Often available sensor data sheets

but calculations based on using

them are less accurate than the

Steinhart-Hart Calculations

calibration factors.

Recommended Steinhart-Hart Cal Parameters

Thermistor type 1		Use the follo defining the	owing opera	parameters for tion of the 3 K
1 Туре	1	Ohm thermis	tor tem	perature sensor
2 Resistance at T0 (ohms)	3000	often used	bv	Vibrating Wire
3 T0 (Celcuis)	25	sensors	~,	
4 Beta	5234	30113013.		
5 Steinhart-Hart 0th order (A)	3.35E-3			
6 Steinhart-Hart 1st order (B)	2.56E-4	(Stein	nart-H	art Factors)
7 Steinhart-Hart 2nd order (C)	2.08E-6			
8 Steinhart-Hart 3rd order (D)	7.30E-8	A	=	3.35E-3
	7.002 0	В	=	2.56E-4
U Up. T Top.	Fig 21	С	=	2.08E-6
	9 – 1	D	=	7.30E-8

The VibWire-108 range of instruments uses the Steinhart-Hart parameters in priority to Beta value temperature calculations once the calibration parameters have been defined.

3000

5234

25

0.0

0.0

0.0

0.0

Thermistor type 1

- 1 Type 2 Resistance at T0 (ohms) 3 T0 (Celcuis)
- 4 Beta
- 5 Steinhart-Hart 0th order (A) 6 Steinhart-Hart 1st order (B)
- 7 Steinhart-Hart 2nd order (C)

U Up. T Top.

Warranty Information

In no event shall Keynes Controls Ltd . be liable for any claim for direct, incidental, or consequential damages arising out of, or in connection with, the sale, manufacture, delivery, or use of any product.

⁸ Steinhart-Hart 3rd order (D)

The information in this document is subject to change without notice. Keynes Controls Ltd . has made a reasonable effort to be sure that the information contained herein is current and accurate as of the date of publication. Keynes Controls Ltd. makes no warranty of any kind with regard to this material, including, but not limited to, its fitness for a particular application. Keynes Controls Ltd will not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

23.0 Modbus Supported Instrument

The VibWire-108 supports Modbus protocol across the RS-485 digital network and acts as a slave unit only.

The layout of the registers used to hold the sensor data values is shown in the tables below.

Modbus version instruments do not operate within Q-LOG and require Modbus client software to operate.



RS-485 Network Connection

The VibWire-108-Modbus has the calibration factors for the sensors installed via the terminal port. See page 22 for details. The same procedure for assigning calibration factors is used

The number and type of sensor inputs to scan is assigned from the keyboard using the menu system. See page 26 for more

The instrument scans automatically at the preset sample period and updates the register values upon completion of the operation. New data values are only transmitted to the Modbus master upon receipt of the Read Register Input Command.

The following Modbus command is used to get data from the

VibWire-108 User Manual V1.11

Fig 22

23.1

details.

23.3

Set-up

throughout the VibWire-108 range.

Scanning The Instrument

1S, 5S, 15S, 1 MIN, 1 HR, 6 HR, 24 HR

0000 0000 0000

Model: VibWire-108-485

23.2 **Data Format**

The Modbus version of the instrument stores data into a series The VibWire-108 scans the Modbus network the automatically of 4 byte registers as shown below. Information is stored as a at a preset time period. floating point 4 byte number. The data is Hex format with the high word being the first 2 bytes and the last being in the next 2 The User can select from scan period of: bytes as shown.

System Information

The last 2 registers in the VibWire-108 are used to check the VibWire-108 data integrity. Register with address 32 increments upon the completion of an instrument scan and is used to show that the Read Input Registers (FC=04) command instrument is still operating.

Register with address 34 increments when the VibWire-108 receives a new Modbus 'Read Input Registers' FC=04 command.

Floating Point Data Value

The tables below show how the registers holding the VibWire-108 data is stored. Address Offset Parameter Description Address Offset Description High order word Parameter Chan-0 Temp 16 17 0 Chan-0 Freq High order word 18 Chan-1 Temp 1 Low order word 19 2 Chan-1 Freq High order word 20 Chan-2 Temp 3 Low order word 21 High order word 4 Chan-2 Freq 22 Chan-3 Temp 5 I ow order word 23 6 Chan-3 Freq High order word 24 Chan-4 Temp 7 Low order word 25 26 Chan-5 Temp 8 Chan-4 Freq High order word 27 9 Low order word 28 Chan-6 Temp 10 Chan-5 Freq High order word 29 Low order word 11 30 Chan-7 Temp 12 Chan-6 Freq High order word 31 13 Low order word 32 Number of Scans 14 Chan-7 Freq High order word 33 15 I ow order word Number of Modbus 34 read attempts 2 Bytes

Low Word

2 Bytes

High Word

Low order word High order word Low order word High order word Low order word High order word I ow order word High order word Low order word High order word

Low order word

35

Modbus over 485 Network Configuration

The images below shows the 485 network for Modbus operations.

Connection Setup		Rea	ad/Write D	Definition			×
Connection Setup Connection Serial Port COM169 G600 Baud B Data bits Even Parity I Stop Bit Port Connect IP Address Port Connect I27.0.0.1 Connect So2 COM16 Connect C	OK Cancel Mode RTU ASCII Response Timeout 1000 [ms] Delay Between Polls 20 [ms] t Timeout [ms]	SI Ad Sa	lave ID: (unction: (ddress: uantity: can Rate: Disable ReadA Disable View Rows © 10 Display: Float AB (1 04 Read In 0 40 1000 w/rite Disabl e on error	put Registers Protocol ad [ms] ed 50 100	: (3x) Idress. E.g. 30	OK Cancel 011 -> 10 Apply ead/Write Once antity alumns ell es (Base 1)

23.4 RS-485 Network Connection & Installation

The VibWire-108-Modbus instrument will connect to any suitable Modbus system supporting RS-485 digital communications. This could be a plant wide SCADA solution or simply a stand-alone system running on a PC or laptop. So long as a suitable communications port is available then the instrument will communicate.

The image below shows how the VibWire-108-Modbus units are connected to the media converter or similar 485 digital communications port. The Modbus operations are transparent over the network. The Keynes Controls USB-485 media converter is shown but any other similar device can be used. The VibWire-108-Modbus instrument operates as a master /slave system where the SCADA system or data recorder is the master and the instrument acts as the slave.





DISP := This option is used to select the type of engineering results that are shown on the 7 segment display.

Per = 1/ Freq = period of oscillation in mSec Freq = XXXX.X in Hz - units changed by terminal Pst = Percentage of range

the percentage of range is used to optimise the settings for the analogue output port in order to achieve the best result

Example. A vibrating wire sensor showing a PSt = 0.1 on the display and settings of LF (Low frequency) of 500 Hz and range of 2000 Hz. This indicates that the results from the sensor is only operating over the lower 10% of the defined range i.e 500 - 700 Hz

For the assigned range above the analogue output port has a scaling of 0 - 2.5V DC using a 16 bit DAC (65536 levels) and will therefore have 0V = 500 Hz & 2.5V = 2500 Hz, so

resolution = (2500-500)/65536 = 0.03 Hz.

However with the instrument adjusted to operate over the range of 500 to 700 Hz as observed above then

0V = 500 Hz & 2.5V = 700 Hz so resolution = 200/65536 = 0.003 Hz PEROD := Sensor Excitation Period

This option defines the update period for the sensor excitation and measurement operation. There is always a trade off between the sensor update rate and the power supply requirements of the instruments.

For fast dynamic changes a fast sensor update rate is required but for slow static measurements only a low update rate is required.

Each instrument requires a unique identification number that is required to locate a specific instrument upon a network. Currently the identifier is an integer of range 0 .. 32.

TRa := Transmission Data Options. (Not used RS485/SDI-12)

To optimise the network bandwidth in order to ensure the maxium number of sensors can be deployed the User is allowed to select the number and type of sensor inputs used on the VibWire-108 for data transmission across a network.

- **1S OT** represents 1 vibrating wire sensors no temperature
- 2S 0T represents 2 vibrating wire sensors no temperature
- 1S 1T represents 1 vibrating wire sensor 1 temp sensor
 2S 2T represents 2 vibrating wire senors 2 temp sensors

Tra.int := Defines the rate of data transmission across a cable free network.

The transceiver fitted inside the VibWire-108 is powered on during the data transmission operation and powered off into a sleep mode between updates.

The faster the data transmission rate the greater the power required to drive the instrument. A compromise is needed to adjust power requirements to the amount of data to be transmitted to guarantee that not readings are lost. (Not used on RS485/SDI-12 networks)

24.1 Keyboard Operations

The keyboard menu system has been designed to be easy to use. Use the menu keys

Move up and down the menu system until the desired parameter is shown in the display. Use the '**Up**' and '**Down**' keys to change the values. Once the new value is selected press the '**Menu Out**' button to store the new value.

25.0 Vibrating Wire Sensors

The VibWire-108 supports most manufacturers sensors and a small summary of these is shown below.



Figure 36

Figure 37





Encardio-rite Vibrating Wire are well tested with the VibWire-108 and are the recommended sensor range for this product range.

Frequency signals i.e. those generated using vibrating wire sensors are particularly suitable for demanding environments such as that often occur within Civil Engineering applications

Vibrating wire sensors ideally suit the harsh civil engineering environment since the signals are capable of long transmission distances without degradation, tolerant of wet wiring conditions and resistant to external electrical noise.

Specifications

Excitation: Voltage Protection: Thermistor: Over-range: Resolution: Accuracy: Thermal Effect: Operating range: Cable: Typical range: Nominal zero value: Auto-resonance Semitron Bi-Polar 230 V 3k Ohms at 25 °C minimum twice pressure 0.025% FSO (minimum) < 0.25% FSO (see Notes) < 0.02% FSO /°C -20 °C +100 °C 2 pair PVC outer sheath. 3500-2200 Hz 3130 Hz

For additional details contact : Encardio-Rite Ltd India

Encardio-rite







Surface strain gauge





Embedment Strain Gauge

Crack meters are available with ranges up to 100 mm

Resolution Accuracy Temperature effect Operating temperature Cable: 2 and 4 Core screened 0.025% of range ± 0.2% of range -0.02mm/°C (typical) -30 to +70°C

0.025% of range ± 0.2% of range

-30 to +70℃

2 Core screened

-0.02mm/℃ (typical)

Joint meters are available with ranges up to 100 mm

Resolution: Accuracy: Temperature effect: Operating temperature: Cable:

Gauge Type: Gauge factor: Measurement range: Resolution: Coil Resistance: Operating temp range: Thermal coefficient:

Effective Gauge length: De-bonded length: Overall length: Standard diameter : Resolution : Strain range: Thermal expansion:

Range: Resolution: Accuracy: Thermal Coefficient: Operating Temp: TSR/5.5/T surface mounting strain gauge. 3.025 x 10-3 micro-strain per Hz squared. > 3000 micro-strain. Generally better than 1 micro-strain. Approximately 100 ohms. -200C to 800C. 11 ppm per 0C.

> 50 mm (nominal) 175 mm 1.39 m 15 mm 0.5 micro-strain 2500 micro-strain 12 ppm /deg C

3,000 micro-strain, set mid-range. 1 micro-strain ± 0.1% FS. 11 ppm / ℃. -20 ℃ to 80 ℃.

The details on this page have been obtained from the sensor manufacturers web sites and are liable for change at any time.

26.0 Main Menu - Terminal Port

Models VibWire-108-SDI12 and VibWire-108-RS485 support a terminal menu system to enable the instrument to be configured.

Refer to page 22 for details of connecting the instrument to a 'Terminal Emulator' program running on a PC.

Ma								
1 System W 2 Thermist 3 Thermist 4 Diagnost 5 Pluck Co 6 Channel 7 Channel 8 Channel 8 Channel B Channel B Channel D Channel U Up. T T	in Menu laintanenu or type ics nutrol 2 3 4 5 6 7 7 0P	2 2 2						
onnected 00:05:40	Auto detect	9600 8-N-1	SCROLL	CAPS	NUM	Capture	Print echo	[]

This is the **Main Menu** that appears on starting the menu system

Quick Menu Guide

- 1. Option 4 'Diagnostics' use menu system available here to start a instrument scan on demand and observe data.
- 5. Option 5 'Pluck Control' restricts sensor ping frequency Used when poor quality sensors with 3rd harmonic oscillations are encountered.

The 'Initial Pluck ' defines the start frequency of the sensor scan.

By default use the automatic sensor excitation '0' as this gives the best result for the majority

The 'Initial Pluck' frequency is a global setting and is of use only then the same model of sensor

27.0 Sensor Excitation Control

The pluck control system built into the VibWire-108 is a useful feature to activate upon observing unusual spikes in what should be steady state data values for sensors that change little over time.

Spikes in the VW Sensor Data

Depending on how well a vibrating wire sensor is made the sensor coil could become damaged, or if the sensor extreme physical shock once it is deployed. Damage to the sensor often means the coil seating has been damaged an the sensor can oscillate at a different harmonic than the designed fundamental frequency.

In order to obtain the correct sensor frequency in the face of oscillations from higher harmonics then the pluck control feature is used.

of sensors

is used on all sensor inputs.

Setting the Pluck Control

Go to the 'Pluck Control' menu as shown in Fig 39 below.

Select the channel to be configured.

Enter the 'Centre Frequency' for the normal operation of the sensor,

The operating frequency for the VW sensor input is now limited to a minimum frequency of $\frac{1}{2}$ of the 'Centre Frequency', and to a maximum of 2 x 'Centre Frequency'. This range removes the third harmonic oscillation which is a common cause of spikes in VW data .



28.0 Firmware Upgrade Facility

Using the Terminal Port menu system

- 1. From the 'Main Menu' select option 1 'Systems Maintenance'
- 2. The following menu will appear -

Firmware Update

Any new firmware is sent out from Keynes Controls technical support only. Only a competent software engineer should undertake this task.

Keynes Controls offer a back to base firmware upgrade service. A small cost is incurred if using this service.

Make sure the latest firmware which is in the form of a text

Example firmware for this documentation is titled 'vw101.txt'

data file, is stored into a suitable location.

🗞 kjgf - HyperTerminal 🗞 kjgf - HyperTerminal		×
<u>File Edit View Call Iransfer Help</u>		
D 📽 🐵 🏂 🕮 🖆 📸		
System Maintanence 1 Upload new firmware 2 Diagnostic Messages 3 Load Defaults U Up. T Top.		
	Fig 40 - System Maintenance Menu	H
<u></u>		

3. Select option 1 '**Upload new firmware**'

🏶 KJYG - HyperTerminal							\mathbf{X}
Eile Edit ⊻iew ⊆all Iransfer Help							
0 🗃 🖉 💲 📽							
System Maintanen 1 Upland New Firmware 2 Diagnostic Messages 3 Load Defaulte U Up. T Top.READYTORCU	ce						
							-
Connected 0:06:05 Auto detect	9600 8-N-1	SCROLL	CAPS	NUM	Capture	Print echo	.::

Once option '1' is selected then the 'Upload new firmware' Window, as shown opposite will appear.

4. Using the HYPER-TERMINAL menu system

Select 'Transfer\Send Text File' option.

KJYG - HyperTerminal	
le Edit Wew Call Transfer Help De Start File Desting Edit	
Syst Cadure Fac Support Cadure Fac 1 Upload New 2 Diagnostic Cadure to Printer 3 Joad Bear Cadure to Printer Up. T Top. READYTORCU	
Fig 41 - Uploading new firmware	
	are
' Burning ' message shows that the firmware has been loaded correct	at y.

Locate and select the new firmware data file.

Fig 41 opposite shows how the 'Hyper-terminal' software appears once the firmware file is selected and data is being sent to the sensor interface.

Fig 42 shows how the 'Hyper-terminal' emulator 'System Maintenance' Window will appear once the firmware has loaded into the sensor interface correctly.

The 'Burning' message has to be displayed to show the new firmware has been loaded correctly.

e E	KJYG-HyperTerminal	×
	Systen Maintanence 1 Upload New Firmware 2 Diagnostic Messages 3 Load Defaults U Up. T Top.REAPYTORCU	<
۶		

Fig 42 - Firmware upload successful

Main Menu



Thermistor Type 1 - Menu

🎝 kjgf - HyperTerminal		- • - ×
🎝 kjgf - HyperTerminal		
Eile Edit View Call Transfer Help		
D 📽 🎯 🎖 🗈 🎦 🗳		
Thermistor type 1		
1 Type 2 Resistance at T0 (ohms) 3 T0 (Celcuis) 4 Beta 5 Steinhart-Hart 0th order (f) 5 Steinhart-Hart 1st order (8) 7 Steinhart-Hart 3rd order (0) 8 Steinhart-Hart 3rd order (C) U Up. T Top.	0 3300.0 25.000 6.00035560 0.00335560 2.06257E-6 2.06227E-6 7.3003E-8	Ŧ

Pluck Control - Menu

🗞 kjgf - HyperTerminal 🌯 kjgf - HyperTerminal	
Eile Edit View Call Transfer Help	
0 2 3 0 3 0	
Pluck Control 1 Initial Pluck 2 GMD Centre Frequency 5 GHZ Centre Frequency 5 GHZ Centre Frequency 7 GHS Centre Frequency 8 GHZ Centre Frequency 9 GHZ Centre Frequency 9 GHZ Centre Frequency 9 GHZ Centre Frequency 0 Up. T Top	0 0 0 0 0 0 0 0 0 0
	Fig 43 - Pluck Control Menu System

By::::::::::::::::::::::::::::::::::::	👌 kjgf - HyperTerminal		
Def is the year is an interest Heip Def is \$ in Def is \$ Channel 2 Image: Interest Processing is the interest of the interest processing is the interest of the interest processing is the interest of the	kjgf - HyperTerminal		
Channel 2 1 Frequency proc 2 Thermistor type 9 3 Cal A 9.0 4 Cal B 1.0000 5 Cal C 0.0 0.0 0.0	<u>Eile Edit View Call Iransfer Help</u>		
Channel 2 1 Frequency proc 0 1 Frequency proc 9 3 0 2 Thermistor type 9 3 0 0 3 Cal A 0.0 0 4 Cal B 1.00000 5 Cal D 0.0 0.0 0 6 6 0 0 0 Cal D 0.0 0.0 0.0 0 0 0 0	D 📽 🎯 🏂 🛍 🎦 🗳		
1 Frequency proc 0 2 Themistor type 9 4 Cal B 1.0000 5 Cal C 0.0 6 Cal D 0.0 0 Cal C 0.0 0 Cal D 0.0 0 J D 0.0 0	Channel 2		
	1 Frequency proc 2 Thermistor type 3 Cal H 4 Cal B 5 Cal C 6 Cal D U Up. T Top	0 9 0.0 1.0000 0.0 0.0	

This is the default menu that appears upon activating the terminal menu system.

Models VibWire-108-SDI12 and VibWire-108-RS485 support 2 individually configured thermistor temperature sensors.

The sensors configuration details are stored as 'Thermistor type 1' and 'Thermistor type 2'.

The menu opposite shows the thermistor type menu.

Use this menu to set the thermistor calculation to Steinhart-Hart or Beta.

The pluck control menu system defines the operating range of a vibrating wire sensor input channels. The sensor excitation frequency will never exceed the limits set here.

The pluck control system prevents 3rd harmonic sensor frequencies from contaminating the measurements due to sensor coil problems.

Each vibrating wire sensor input can be individually configured.

 Range =
 (1/2 x Centre Freq)
 Centre Freq - (2 x Centre Freq)

 Low Frequency
 Centre Frequency
 High Frequency

The image opposite shows the default Vibrating wire sensor input channel configuration menu system.

Option 1 'Frequency' - Hz , Hz² or SI units

Repeat for each sensor input channel.

Case Study - Bore Hole Pump Control

The following examples detail the use the VibWire-108 range of For applications where a control system is required to let a set instruments within water flow and control systems. Two amount of water to flow and to maintain a record of events then examples are shown. The first example demonstrates the a control system similar to that shown below will be required. VibWire-108 as a stand-alone control system used to monitor Apart from maintaining details of flow the control unit can also bore hole water and the second example, shows how the act as an alarm system and shut of water flow should the VibWire-108 is used with a data logger to report on and control defined limited be exceeded. the water flow in a v-notch weir.

Borehole Water Level Control

The VibWire-108 can be configured to operate as a stand-alone available for this calculation but they depend upon the water control system when used with Vibrating wire piezometers and height at different positions within the weir. the transistor output ports.

Fig 43 shows a sealed bore hole with 2 piezometers deployed. water height and are connected to the VibWire-108. The lower level sensor reports the water level height. The top sensor acts as a reference and reports the internal bore hole grammable logger, such as the Keynes Controls AguaLOG pressure only.

The water level height measurement can be in error if there is a build up of gas pressure in the hole and it is to correct for gas pressure changes that the reference is fitted.



The VibWire-108 is configured to activate the digital output when the water height exceeds the higher alarm levels. Upon detection of a water level exceeding the pre-determined upper alarm level then the transistor output is activated to switch a relay.

The relay switching can be used to activate the ignition circuit on the pump if it is powered by a petrol engine or switch the power supply if an electric pump is being used.

Summary

As long as the Peizometer frequency settings are defined correctly and a suitable hysteresis level is set, then the VibWire-108 can run indefinitely to control a pump to maintain the water level within a bore hole below a pre-set level.



Stand-alone Weir Control System with Data Logger

Consider the example of the V-notch weir below. A dedicated complex formula is required to calculate the discharge over the bottom of the notch. There are a number of different formula

A series of vibrating wire piezometers are used to monitor the VibWire-108 communicates the water height details to a prosystems. The logger is used to calculate the discharge and to record the results.



Fig 42 shows the basic control system showing the VW-108 connected to a logger across a digital network.

The instantaneous pressure levels are monitored by the logger and used within the Vnotch discharge formula.

The logger can be used to produce a permanent result record that can be downloaded by an operator and also to trigger the transistor outputs within the VibWire-108 when used for alarm level control.

Fig 45 shows the v-notch weir 7777 profile. The water height is measured accurately using a precision vibrating wire sensor and these types of sensors are very accurate and stable over long periods of time.





vibrating wire sensor data and convert it into engineering units.

The various sensors forming the weir monitoring system are combined within the AquaLOG to determine the correct flow rate. The Alarm operations of the AquaLOG can then be used to inform an operator should pre-set conditions be exceeded.

Vibrating Wire Sensor Applications

civil engineering applications.

The details below show just a few of the applications for which A lot of vibrating wire sensor applications are undertaken with vibrating wire sensors are used. The best results from the the sensors buried into the structure such as within concrete or instrumentation are only achieved when a good understanding within a bore hole. The vibrating wire sensors maintain very for the overall structural behaviour is available. The examples accurate results without need for re-calibration over many years demonstrate the common uses of structural monitoring within and generally more reliable than other forms of sensor, less prone to failure in damp conditions and relatively easy to install.





The images opposite show sister bar strain gauges and examples of their installation and use. The sister bar strain gauge comprises of two lengths of ribbed rebar welded to a central gauge section. The central gauge section has a miniature stainless steel, vibrating wire strain gauge element, fitted along the longitudinal axis the of gauge. Incorporated within the gauge section are two coils for excitation and output of the vibrating wire strain gauge element. This type of gauge is particularly rugged and reliable. They are particularly applicable for strain measurements in mass concrete pours where placing of concrete is remote and uncontrolled such as typically occurs in diaphragm walls or deep piles.

The sister bars can be connected to the VibWire-108 instruments.



The embedment or dumbbell gauge (See below) is designed to measure strain in concrete. This vibrating wire strain gauge is typically tied to a reinforcing cage. It is usual to install them in arrays of three or four gauges at several depths horizons within the structure.



VibWire-108/101



Temperature Measurement - Using the Steinhart-Hart Equation for temperature calculation

Introduction

The following technical note shows how to obtain the thermistor resistance values for a vibrating wire temperature sensor connected to the temperature input of the VibWire-101 or VibWire-108 interfaces.

The VibWire-108/101 vibrating wire sensor interface supports the 4 wire sensor interface. This page demonstrates how the temperature calculation is carried out by the VibWire instruments. Page 23 shows the menu configuration options for the temperature sensorts.

The results that can be achieved will depend on how well the thermistor calibration factors have been assigned.

The circuit below shows the VibWire-108 temperature input with pull-up resistor completion

The VibWire-101 and 108 models use 2.4 V excitation for the sensor thermistor.



V_{therm} = Voltage across thermistor

V_R = Voltage across pull up resistor

Example. A VibWire-108/101 provides an output temperature value of 1086 mV then

 $I_{therm} = (2.4 - V_{therm}) / 3300 \text{ where } 3300 = \text{pull-up resistor value} \text{ where } V_{therm} = 1.086 \text{ V}$ therefore $I_{therm} = (\text{Excitation volt - V}_{therm}) / 3300_{(Pull-up Resistor)} = (2.4 - 1.086) / 3300 = 1.414 / 3300 = 0.398 \text{ mA}$

using Ohms Law

Note 1086 mV = 1.086 Volts

The Resistance of the Thermistor is calculated

 $R_{therm} = V_{therm} / I_{therm} = 1.086 / 0.000398 = 2727.4 Ohm$

Now 2727.4 ohms is the resistance of the thermistor at the at temp (T)

Temperature Conversion

The thermistor resistance value is converted to temperature using the Steinhart-Hart Equation.

$$T = \underbrace{1}_{C_1 + C_2 . \text{ In} \setminus R_{\text{therm}} + C_3 (\text{In} R_{\text{therm}})^3} \text{ where } T = \text{absolute temperature in Kelvin } R_{\text{therm}} \text{ in Ohms}$$

Conversion to Deg C is

$$T(C) = \frac{1}{C_1 + C_2 . \text{ In } R_{\text{therm}} + C_3 (\text{In} R_{\text{therm}})^3} - 273.15$$

The sensor data sheet will show for the thermistor a calibration equation similar to that below. The values for the parameter C_1 , C_2 , & C_3 will be listed.

 $(1/T) = C_1 + C_2 Ln(R_{therm}) + C_3 Ln(R_{therm})^3 - 273.15$

Example

In Vibrating Wire sensors is the 44005RC Precision Epoxy NTC Thermistor is commonly used for temperature monitoring applications.

The data sheet for this product can be downloaded at

http://www.aquabat.net/downloads/1350009-2.pdf - The thermistor data sheet is valid to 11/12/2013 refer to the manufactures data sheet for the latest information.

An example Excel spreadsheet that demonstrates the temperature calculations can be downloaded at

http://www.aquabat.net/downloads/ThermistorWorksheet.xls

Example

The VibWire-101 is can be set to give ratiometric or mV temperature values from the built in thermistor of a vibrating wire sensor. depending upon the sensor configuration. Ratiometric values are calculated between the 3300 Ohm pull up resistor and thermistor resistance and is value between 0 - 1. The Vibwire-101 has returned a value of 0.663 from the thermistor.

In the spreadsheet below the VW-101 gives a temperature value (Ratiometric) of 0.663. The constants A, B and C are from the calibration data sheet. The spreadsheet below shows the temperature to be 7 Deg C,



Voltage ratio 0.663 Input Excitation (Ohm) 3300 Fixed Thermistor resistance 6905 Calculated Thermistor R0 3000 Thermistor property А 1.41E-03 Thermistor property В 2.37E-04 Thermistor property С 1.02E-07 Thermistor property Inv Temperature 3.57E-03 Temperature (Celsius) 7.0 Calculated value Temperature Steinhart-Hart Calibration value Parameters obtained from calibration data sheet.

Calculation of temperature based on voltage ratio

Thermistor Temperature Calculation using the thermistor Beta Value

An NTC (Negative Temperature Coefficient) Thermistor is a passive electrical component whose resistance varies inversely with temperature. It is often used as a temperature sensor inside vibrating wire sensors.

The relationship between resistance and temperature can be described with the 'beta' formula.

The VibWire-108 range of sensor interfaces offers the beta value temperature calculation as an option within the thermistor setup. The Beta value thermistor calculation is a simplified version of the Steinhart-Hart equation that is most often used in temperature measurement solutions.

For applications when the Steinhart-Hart calibration factors are not known then the sensor 'Beta' value can often be found on the thermistor manufactures data sheet.

The example below demonstrates how the VibWire-108 calculates temperature

The temperature calculations are undertaken internally within the VibWire-108 and are not yet currently part of the Q-Log software.

Refer to Page 23 for assigning the Beta value into the

$$T = \frac{B}{\ln\left(\frac{Re}{T_0}\frac{B}{R_0}\right)}$$

where T = temp in units K (Kelvin)

B = Thermistor Beta value

T₀ = Temperature at 25 ° C

R₀ = Resistance at 25 ° C

and to convert Deg K to Deg C then $T_o = T_k - 273$



at 25 Deg C

Encardio-rite Electronics Pvt. Ltd.

A-7 Industrial Estate, Talkatora Road, Lucknow, UP-226011 India

E-mail: geotech@encardio.com, lko@encardio.com; Website: www.encardio.com Tel. +91 (522) 2661039/40/41/42 Fax +91 (522) 2662403



Appendix - C

TEST CERTIFICATE DWT Traceable to standard no. : J082301 T8F 281 TC

Customer	:					
P.O. No.	:					
Instrument	:	V W Piezome	eter		Date : 02	2.02.2012
Serial number	:	XXXXX			Temperature : 1	9°C
Capacity	:	350 kPa			Atm. Pressure : 1	00 kPa
Input		Observed v	alue	Average	End Point	Poly
pressure	Up1	Down	Up2		Fit	Fit
(kPa)	(Digit)	(Digit)	(Digit)	(Digit)	(kPa)	(kPa)
0.0	6555.9	6556.9	6556.9	6556.4	0.0	0.3
70.0	6312.4	6312.6	6312.4	6312.4	69.3	69.5
140.0	6064.0	6064.3	6063.1	6063.5	139.9	140.1
210.0	5817.1	5818.4	5816.2	5816.7	210.0	210.1
280.0	5569.8	5570.7	5568.0	5568.9	280.3	280.3
350.0	5323.3	5323.3	5323.7	5323.5	350.0	349.8
Digit	:	f ²/1000		- '^'``		- · ·
Linear gage factor (G)	:	2.8388E-01	kPa/digit			
		(Use gage fa	actor with mir	ius sign with o	ur read out unit Mo	odel : EDI-51V)
Thermal factor(K)	:	-0.087	kPa/°C			
Polynomial constants	:					
	$\langle A \rangle$	A= -2.2253E-07) (В	= -2.8085E-01) (C= 1	1.8512E+03
Pressure "P" is calcula	ted with th	e following equ	ation:			
Linear	:	P(kPa) =G(F	R0-R1)+K(T1-	-T0)-(S1-S0)		
Polynomial	:	P(kPa) = A(l	$(R1)^{2} + B(R1)$	+ C+K(T1-T0)-	-(S1-S0)	
		R1 = curre	nt reading & I	R0 is initial rea	ding in digit.	
		S1 and T1 =	current atmo	spheric press	ure(kPa) and temp	erature (°C)
Readings at the time o	f shipment	:	Date			
f	:		Hz	The te	erms K(T1-T0) are	e the temperature
f ²	:		Digit	compe	ensation terms	tor this sensor.
Temperature	:		°C	Taura		teal weather and
Thermistor	:		Ohm	I empe	erature compensa	ted readings only
Atm.pressure	:		kPa	WORK I	the thermsitor op	eration is defined.
Coil resistance	:		Ohm			

(Zero conditions in the field must be established by recording the reading R0 (digit) along with temperature T0 (°C)and atmospheric pressure S0 (kPa) at the time of installation. If polynomial constants are used, determine value of 'C' as per § 6.2 of user's manual.)

Main Menu	Frequency proc		The example above is for a VW Piezometer and so the engineering units calculations vary
1 System Maintenance 2 Thermistor type 1 3 Thermistor type 2 4 Diagnostics 5 Channel 0	0 = Hz 1 = Digits (Hz ² where SI Units is by Qu	²) 2 = SI Units uadratic Expansion	For a Piezometer the local barometric levels are taken into consideration.
6 Channel 1 7 Channel 2 8 Channel 3 9 Channel 4 A Channel 5 B Channel 6 C Channel 7 U Up. T Top.	Channel 0 1 Frequency proc 2 Thermistor type 3 Cal A 4 Cal B 5 Cal C 6 Cal D	1 -2.2253E-07 -2.8085E-01 1.8512E+03 -0.087	The engineering units for this example is K Pa The term (S1-S0) is a constant offset that allows for local atmospheric conditions and be taken from a barometer module such as models Barom-SDI12 or Barom-485 . The VW sensor units have to be set to 'Digits'

U Up. T Top.

that is $Hz^2/1000$.

VibWire-108 User Manual V1.11

VIBRATING WIRE INSTRUMENTS CALIBRATION CERTIFICATE

Instrument Type : Displacement Transducer	Serial No.	: 012453
Instrument Range: 0.00 to 50.0 mm	Calibration Date.	: 14th March 2014
Gauge Factors in mm	Ambient Temp.	: 23 Deg C
Period Gauge Factor (K : 92.1053900	Barometric Pressure	: 1015 mb
Linear Gauge Factor (G) : (mm/digit) -0.0092090	Calibrator Personnel	: Ian Thomas
Polynomial Gauge Factor A: 0.000000024979750	Calibration Equipment	th scale
Polynomial Gauge Factor B: 0.0089750451	VibWire-108 sensor in	terface
Polynomial Gauge Factor C: 28.976750	Regression Zero:	3185.7

Reading (Period)	Digits F ² /1000	Calculated (Linear)	Error %FS (Linear)	Linear Increment	Applied (mm)	Calculated (Polynomial)	Error %FS (Polynomial)
5610.9	3176.4	-0.088	-0.18	0.0	0.00	0.023	0.05
5182.9	3722.6	4.943	-0.11	546.2	5.00	4.987	-0.03
4840.0	4268.8	9.974	-0.05	546.2	10.00	9.966	-0.07
4555.8	4818.0	15.032	0.06	549.2	15.00	14.988	-0.02
4316.6	5366.8	20.087	0.17	548.8	20.00	20.021	0.04
4112.2	5913.5	25.123	0.25	546.7	25.00	25.049	0.10
3937.9	6448.8	30.053	0.11	535.3	30.00	29.987	-0.03
3782.8	6988.5	35.024	0.05	539.7	35.00	34.981	-0.04
3643.9	7531.2	40.023	0.05	542.7	40.00	40.017	0.03
3521.8	8062.5	44.917	-0.17	531.3	45.00	44.961	-0.08
3409.0	8604.8	49.912	-0.18	542.3	50.00	50.022	0.04

Formulae:	Linear	$E = G(R_0 - R_1)$
	Polynomial	$E = AR_{1}^{2} + BR_{1} + C$
	Offset	$C = -(AR_0^2 + BR_0)$

These equations give displacement only without any temperature compensation.

Main Menu

1 System Maintenance

2 Thermistor type 1

3 Thermistor type 2

4 Diagnostics 5 Channel 0 6 Channel 1

7 Channel 2

8 Channel 3

9 Channel 4

A Channel 5

B Channel 6

C Channel 7

U Up. T Top.

Frequency proc 0 = Hz 1 = Digits (Hz²) 2 = SI Units where SI Units is by Quadratic Expansion

Channel 0 1 Frequency proc 2 Thermistor type 3 Cal A 4 Cal B

5 Cal C

6 Cal D

U Up. T Top.

1 1 0.000000024979750 0.0089750451 28.976750 -2.3069E-3 Where R_0 = is the initial zero reading at the time of installation.

 R_1 = current reading Units Hz²/1000 (Digits)

Displacement calculations within Q-Log

Q-Log is the free data display and recording software from Keynes Controls Ltd.

Linear

 $E = G(R_0 - R_1)$ $E = G.R_0 - G.R_1$

where G.R₀ is a constant term.

Enter this formula into Q-log and the results will be in mm.

See

http://www.aquabat.net/QLOGFree/qlogv2.html



CALIBRATION DATA VIBRATING WIRE PIEZOMETER

09-13489 Serial No MAXIMUM PRESSURE 10 MPa MAXIMUM OVERLOAD PRESSURE 20 MPa BAROMETRIC PRESSURE AT CALIBRATION 1012 mBar FACTORY ZERO READING 9061.304 Hz^2/1000 CABLE RESISTANCE (TWO WAY) 51.25 Ohms PRESSURE CHANNEL COEFFICIENTS P4= -1.08331E-15 P3= 0.00000E+00 P2= -2.21240E-05 P1= 0.00000E+00 P0= FROM TEMP CH **TEMPERATURE CHANNEL COEFFICIENTS** C4=8.50983E-14 C3= -2.05402E-09 C2= 1.84317E-05 C1= -8.09166E-02 C0= 2.10758E+04 SWEEP FREQUENCY MIN 1978 Hz SWEEP FREQUENCY MAX Hz 3311 **TEMPERATURE STEINHART-HART** X^3 (B) 1.01900E-07 X^2 (D) 2.36900E-04 X^0 (E) 1.40510E-03 **CYCLES** 250 SETTLING TIME (mSec) 150 **Thermistor Configuration Frequency Input Configuration** Channel 0 **Thermistor type 1** Output in Digits 1 Frequency proc Freq²/1000 1 Type 2 Resistance at T0 (ohms) (1)2 Thermistor type 3000 1 0.0 3 T0 (Celcuis) 3 Cal A 4 Cal B -2.21240E-05

 3 T0 (Celcuis)
 25

 4 Beta
 5234

 5 Steinhart-Hart 0th order (A)
 3.35E-3

 6 Steinhart-Hart 1st order (B)
 2.56E-4

 7 Steinhart-Hart 2nd order (C)
 2.08E-6

 8 Steinhart-Hart 3rd order (D)
 7.30E-8

U Up. T Top.

Common VW Sensor Thermistor Part Numbers

5 Cal C

6 Cal D

U Up. T Top.

YSI 44005 Vishay 1C 3001 B3 RS Part no: 151-215 The part numbers are for 3 K Ohm thermistor commonly used by most different manufacturers.

-1.08331E-15

The sensors give 3 K Ohm resistance at 25 Deg C.

0.0

Dimensions of the VibWire-108 Back Mounting Panel

The image below shows the dimensions of the back mounting panel for the VibWire-108 range of vibrating wire sensor interfaces.



Page No.	Description
2	Introduction
	Hardware Options Dynamic Measurements Configuration, SI Units
3	Features Terminal Port, Q-LOG
4	Index
5	Care & Maintenance Default Factory Settings
6	Front Panel
7	SDI-12 Serial Network Connection SDI-12 Network Operation Multi-instrument Solutions
8	PC Data Acquisition System, Multi-instrument Data Logger Solution RS-485 Serial Network Connection Earth Connection PC Data Acquisition - RS-485 Solution
9	Modbus RS-485 Network Connection
10	Technical Specifications
11	System 1est - Power on sequence VibWire-108 Digital Communications Command Structure & Operations across an RS-485 network Timing Constraints RS-485 / SDI-12 Networks
12	RS-485/SDI-12 Command Format Assigning Data Transmission over the SDI-12 or RS-485 network Starting Data Acquisition Operation on a SDI-12 Network
13	Modbus version instrument Instrument Identifier Start Measurement Commands
	Advice on the choice of the measurement instructions. Possible Network Problems
14 15	SDI-12 Supported Commands RS-485 Supported Commands
16	Examples Of Using RS-485/SDI-12 Commands Changing the instrument ID number ID Number Query
17	Start measurements for instruments on a network Start measurements using the concurrent command Read data values from the VibWire-108
18	Connection to an Analogue Data Acquisition System Theory of operation,
19	Optimising the Analogue Output Settings
20	Real-time Frequency Display Assigning the real-time display to a VW sensor input
21	Sensor Problems Vibrating Wire Sensor Installation Common Earth Points
22	Lightning Protection Terminal Port Terminal Port Operation
23	Terminal Port Menu System Selecting the VW Sensor Output Units
24	Modbus Supported Instrument, Modbus Registers Scanning the instrument
25	RS-485 / Modbus Network Connection & Installation
26	Keyboard Operations
27 28	Vibrating Wire Sensors Main Menu - Terminal Port
	Spikes in the VW Sensor Data Pluck control worked example
29 30 31-31	Firmware upgrade facility Terminal Port Menu Screens Case Study
Appendix A	Temperature measurements - Thermistor Steinhart-Hart Thermistor worked example
Appendix B	Thermistor Temperature Calculation using the thermistor Beta Value
Appendix C	Sample VW Piezometer Calibration Data Sheet & Menu System Configuration details Sample VW Displacement sensor Calibration Data Sheet & Menu System Configuration details Sample Sigra Pressure Sensor - Steinhart-Hart Default Cal Factors used.
39	Pressure calibration factors for SI Units assigned. Physical Dimensions VW108 series instrument mounting plate